
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007325
7:12 07/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:41
STERILIZING 4:43
136.4 C 222kPa 4:43
136.1 C 221kPa 5:13
136.3 C 223kPa 5:43
136.4 C 224kPa 6:13
135.9 C 219kPa 6:43
136.3 C 222kPa 7:13
136.1 C 221kPa 7:43
136.2 C 222kPa 8:13
136.2 C 222kPa 8:13
VENTING 8:14
AIR DRYING 9:09
CYCLE COMPLETE 39:09

Digital Signature #
C5BA32EC5F5ADED4
