
Midmark M11 - v1.0.0
Total Cycles: 3598

Name: RITTER 1
Sterilizer ID: V2178198
08/01/2023 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.1 F 0.0
2:00 195.3 F 0.9
4:00 195.6 F 5.4
6:00 217.8 F 11.1
8:00 224.9 F 6.3
10:00 237.7 F 11.8
12:00 236.5 F 9.3
14:00 239.4 F 9.1
16:00 248.9 F 14.7
18:00 261.1 F 21.6
20:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.3 F 28.8
2:00 271.9 F 28.8
3:00 272.2 F 28.7
4:00 272.0 F 28.8

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.9 F 0.0
5:00 214.2 F 0.0
10:00 230.2 F 0.0
15:00 229.4 F 0.0
20:00 228.4 F 0.0
25:00 227.7 F 0.0
30:00 231.6 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 20:27
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:55
CYCLE COMPLETE

08/01/2023 10:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____