
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000161
16:04 17/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:38
STERILIZING 3:52
136.5 C 220kPa 3:52
136.4 C 224kPa 4:22
136.4 C 224kPa 4:52
136.2 C 219kPa 5:22
136.2 C 220kPa 5:52
136.4 C 222kPa 6:22
136.4 C 223kPa 6:52
136.3 C 223kPa 7:22
136.3 C 223kPa 7:22
VENTING 7:23
AIR DRYING 8:16
CYCLE COMPLETE 8:21

Digital Signature #
DAE929228B398AF6
