
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007285
11:22 29/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:31
STERILIZING 3:34
136.4 C 220kPa 3:34
136.3 C 223kPa 4:04
136.4 C 223kPa 4:34
136.4 C 224kPa 5:04
136.4 C 222kPa 5:34
136.3 C 224kPa 6:04
136.4 C 223kPa 6:34
136.4 C 222kPa 7:04
136.4 C 222kPa 7:04
VENTING 7:04
AIR DRYING 7:54
CYCLE COMPLETE 29:31

Digital Signature #
D09E8F30956DF37F
