
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000127
10:29 08/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:46
PRESSURIZING 4:47
STERILIZING 6:21
136.4 C 224kPa 6:21
136.4 C 225kPa 6:51
136.3 C 225kPa 7:21
136.2 C 222kPa 7:51
136.3 C 224kPa 8:21
136.2 C 223kPa 8:51
136.2 C 222kPa 9:21
136.2 C 223kPa 9:51
136.2 C 223kPa 9:51
VENTING 9:51
AIR DRYING 10:55
CYCLE COMPLETE 11:02

Digital Signature #
54FF1B994DAECB02
