
Midmark M11 - v1.0.5
Total Cycles: 436

Name: RITTER2
Sterilizer ID: V2313228
10/12/2021 07:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 79.7 F 0.0
4:00 81.5 F 0.1
6:00 86.3 F 0.1
8:00 120.0 F 0.5
10:00 156.5 F 3.7
12:00 178.9 F 7.1
14:00 197.4 F 10.7
16:00 209.3 F 6.6
18:00 221.7 F 8.1
20:00 233.1 F 12.4
22:00 236.4 F 9.6
24:00 238.4 F 9.9
26:00 247.2 F 14.1
28:00 258.2 F 20.2
30:00 268.6 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.6 F 29.2
2:00 271.4 F 29.0
3:00 272.2 F 29.0
4:00 271.6 F 29.2

Min 271.1 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 243.2 F 0.0
10:00 227.1 F 0.0
15:00 239.9 F 0.0
20:00 181.6 F 0.0
25:00 170.9 F 0.0
30:00 158.2 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 31:11
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:09:22
CYCLE COMPLETE

10/12/2021 08:55 AM
ACCEPTED: X
REJECTED:
Operator: U01