

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002121  
9:09 04/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:48  
PRESSURIZING 2:55  
STERILIZING 4:09  
136.0 C 220kPa 4:09  
136.1 C 222kPa 4:39  
136.4 C 225kPa 5:09  
136.4 C 221kPa 5:39  
136.2 C 223kPa 6:09  
136.4 C 224kPa 6:39  
136.4 C 224kPa 7:09  
136.4 C 224kPa 7:39  
136.4 C 224kPa 7:39  
VENTING 7:39  
AIR DRYING 8:34  
CYCLE COMPLETE 38:34  
-----

Digital Signature #  
7B88BCC480C83686  
-----