
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000352
9:50 20/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:12
STERILIZING 3:14
136.0 C 227kPa 3:14
136.3 C 233kPa 3:44
136.1 C 230kPa 4:14
136.4 C 233kPa 4:44
136.4 C 233kPa 5:14
136.2 C 232kPa 5:44
136.4 C 233kPa 6:14
136.4 C 233kPa 6:44
136.4 C 233kPa 6:44
VENTING 6:44
AIR DRYING 7:35
CYCLE COMPLETE 12:33

Digital Signature #
8F2C9C67AFD00D5E
