

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001302  
11:30 01/03/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:29  
PRESSURIZING 3:30  
STERILIZING 4:32  
136.0 C 221kPa 4:32  
136.3 C 225kPa 5:02  
136.3 C 224kPa 5:32  
136.2 C 224kPa 6:02  
136.4 C 225kPa 6:32  
136.5 C 225kPa 7:02  
136.4 C 224kPa 7:32  
136.4 C 225kPa 8:02  
136.4 C 225kPa 8:02  
VENTING 8:02  
AIR DRYING 8:58  
CYCLE COMPLETE 8:59

-----  
Digital Signature #  
AD0435EF1C540BB7  
-----