
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002329
8:26 29/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:49
PRESSURIZING 0:57
STERILIZING 2:05
136.4 C 224kPa 2:05
136.1 C 221kPa 2:35
136.3 C 222kPa 3:05
136.1 C 221kPa 3:35
136.2 C 222kPa 4:05
136.4 C 224kPa 4:35
136.2 C 223kPa 5:05
136.4 C 223kPa 5:35
136.4 C 223kPa 5:35
VENTING 5:35
AIR DRYING 6:25
CYCLE COMPLETE 36:25

Digital Signature #
C026B865762DA93F
