
Midmark M11 - v1.0.0
Total Cycles: 4351

Name: RITTER 1
Sterilizer ID: V2178198
03/01/2024 10:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.4 F 0.0
2:00 143.0 F 1.5
4:00 174.8 F 3.8
6:00 179.4 F 7.3
8:00 197.5 F 11.1
10:00 216.3 F 3.7
12:00 221.7 F 6.8
14:00 232.7 F 10.7
16:00 232.2 F 7.4
18:00 241.4 F 11.4
20:00 237.3 F 8.9
22:00 249.6 F 14.7
24:00 259.8 F 20.4
26:00 269.4 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 273.1 F 28.8
2:00 272.1 F 28.9
3:00 271.4 F 28.4
4:00 272.4 F 28.5

Min 271.0 F 28.0
Max 273.1 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.3 F 0.0
5:00 200.7 F 0.0
10:00 195.7 F 0.0
15:00 192.2 F 0.0
20:00 190.1 F 0.0
25:00 189.3 F 0.0
30:00 188.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 27:07
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:05:49
CYCLE COMPLETE

03/01/2024 11:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____