
Midmark M11 - v1.0.0
Total Cycles: 1798

Name: RITTER 1
Sterilizer ID: V2178198
04/06/2022 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.0 F 0.0
2:00 181.3 F 2.7
4:00 185.5 F 6.7
6:00 205.4 F 11.6
8:00 222.3 F 5.5
10:00 231.6 F 10.7
12:00 232.0 F 7.0
14:00 245.1 F 13.5
16:00 243.9 F 12.1
18:00 256.5 F 18.8
20:00 267.7 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.7 F 28.9
2:00 271.4 F 28.4
3:00 272.6 F 28.9
4:00 271.5 F 28.2

Min 271.3 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 221.3 F 0.0
10:00 222.6 F 0.0
15:00 230.2 F 0.0
20:00 229.4 F 0.0
25:00 229.7 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 21:18
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:59:42
CYCLE COMPLETE

04/06/2022 11:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____