

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2506

Name: RITTER 1  
Sterilizer ID: V2178198  
11/02/2022 06:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
2:00 196.9 F 4.5  
4:00 215.9 F 9.3  
6:00 226.1 F 7.7  
8:00 230.9 F 10.1  
10:00 232.7 F 7.4  
12:00 245.7 F 13.2  
14:00 243.6 F 11.8  
16:00 255.9 F 18.5  
18:00 267.2 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.0 F 28.5  
2:00 272.0 F 28.9  
3:00 271.2 F 28.6  
4:00 272.0 F 28.7  
  
Min 271.2 F 28.3  
Max 272.4 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.8 F 0.0  
5:00 238.5 F 0.0  
10:00 232.7 F 0.0  
15:00 230.5 F 0.0  
20:00 229.1 F 0.0  
25:00 229.6 F 0.0  
30:00 230.1 F 0.0  
  
Drying Complete

FILLING: 1:49  
HEATING: 19:17  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:15  
CYCLE COMPLETE

11/03/2022 06:55 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02