

-----  
Midmark M11 - v1.0.5  
Total Cycles: 217

Name: RITTER2  
Sterilizer ID: V2313228  
07/15/2021 01:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.6 F 0.0  
2:00 185.9 F 3.6  
4:00 203.9 F 8.6  
6:00 220.0 F 12.3  
8:00 231.9 F 8.8  
10:00 240.1 F 11.5  
12:00 242.7 F 12.3  
14:00 242.4 F 11.7  
16:00 254.8 F 18.1  
18:00 265.5 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.5 F 29.2  
2:00 271.8 F 29.3  
3:00 272.2 F 29.0  
4:00 271.6 F 29.2  
  
Min 271.3 F 28.7  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.4 F 0.0  
5:00 236.4 F 0.0  
10:00 224.1 F 0.0  
15:00 210.2 F 0.0  
20:00 200.5 F 0.0  
25:00 192.7 F 0.0  
30:00 172.8 F 0.0  
Drying Complete

FILLING: 2:15  
HEATING: 19:46  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 00:59:26  
CYCLE COMPLETE

07/15/2021 02:06 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02