
Midmark M11 - v1.0.0
Total Cycles: 4416

Name: RITTER 1
Sterilizer ID: V2178198
03/28/2024 10:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.4 F 0.0
2:00 185.6 F 2.7
4:00 188.8 F 6.6
6:00 204.1 F 11.8
8:00 216.4 F 5.5
10:00 231.6 F 10.6
12:00 234.1 F 7.1
14:00 243.7 F 13.1
16:00 242.7 F 11.7
18:00 255.5 F 18.2
20:00 266.7 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 29.1
2:00 272.5 F 28.8
3:00 271.8 F 28.3
4:00 271.4 F 28.4

Min 271.3 F 28.1
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 221.5 F 0.0
10:00 230.1 F 0.0
15:00 226.8 F 0.0
20:00 227.5 F 0.0
25:00 229.2 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 21:38
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:00:19
CYCLE COMPLETE

03/28/2024 11:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____