

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002262  
13:35 14/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:42  
STERILIZING 3:49  
136.0 C 221kPa 3:49  
136.4 C 224kPa 4:19  
136.1 C 222kPa 4:49  
136.3 C 225kPa 5:19  
136.1 C 220kPa 5:49  
136.4 C 223kPa 6:19  
136.4 C 225kPa 6:49  
136.4 C 223kPa 7:19  
136.4 C 223kPa 7:19  
VENTING 7:19  
AIR DRYING 8:11  
CYCLE COMPLETE 38:11  
-----

Digital Signature #  
4CCFB23E9979FC72  
-----