
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002961
12:14 13/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:30
STERILIZING 3:30
136.4 C 220kPa 3:30
136.4 C 220kPa 4:00
136.3 C 222kPa 4:30
136.0 C 221kPa 5:00
136.3 C 223kPa 5:30
136.4 C 224kPa 6:00
136.2 C 223kPa 6:30
136.3 C 223kPa 7:00
136.3 C 223kPa 7:00
VENTING 7:00
AIR DRYING 7:54
CYCLE COMPLETE 13:06

Digital Signature #
BF93070FE48CCBFB
