
Midmark M11 - v1.0.0
Total Cycles: 2710

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2022 01:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.2 F 0.0
2:00 175.5 F 3.4
4:00 191.0 F 7.6
6:00 212.4 F 12.8
8:00 220.3 F 7.2
10:00 234.7 F 12.6
12:00 236.3 F 9.8
14:00 236.4 F 8.4
16:00 249.7 F 15.1
18:00 262.0 F 22.1
20:00 272.9 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.9 F 29.3
2:00 272.1 F 28.6
3:00 272.1 F 29.0
4:00 272.8 F 29.3

Min 271.6 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 215.9 F 0.0
10:00 234.8 F 0.0
15:00 237.7 F 0.0
20:00 227.1 F 0.0
25:00 223.8 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 20:24
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:60
CYCLE COMPLETE

12/21/2022 02:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____