
Midmark M11 - v1.0.0
Total Cycles: 2473

Name: RITTER 1
Sterilizer ID: V2178198
10/21/2022 03:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.4 F 0.0
2:00 193.1 F 4.3
4:00 203.8 F 8.7
6:00 223.4 F 7.9
8:00 231.1 F 8.7
10:00 240.7 F 11.2
12:00 240.7 F 11.6
14:00 239.4 F 10.2
16:00 252.9 F 17.0
18:00 264.6 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.6
1:00 272.2 F 29.2
2:00 272.1 F 28.8
3:00 271.6 F 28.8
4:00 272.5 F 29.2

Min 271.4 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 238.7 F 0.0
10:00 228.1 F 0.0
15:00 230.8 F 0.0
20:00 228.6 F 0.0
25:00 230.8 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 19:46
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

10/21/2022 04:37 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____