
Midmark M11 - v1.0.5
Total Cycles: 657

Name: RITTER2
Sterilizer ID: V2313228
12/21/2021 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.4 F 0.0
2:00 126.1 F 1.1
4:00 140.7 F 3.5
6:00 167.5 F 7.2
8:00 213.2 F 12.0
10:00 214.3 F 6.1
12:00 226.5 F 10.9
14:00 228.0 F 6.9
16:00 239.5 F 12.8
18:00 240.0 F 11.2
20:00 252.1 F 17.3
22:00 262.7 F 23.7
24:00 273.0 F 30.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.5
1:00 272.0 F 30.2
2:00 272.9 F 30.3
3:00 273.0 F 30.3
4:00 271.8 F 30.1

Min 271.0 F 29.4
Max 273.5 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 230.6 F 0.0
10:00 223.6 F 0.0
15:00 227.9 F 0.0
20:00 215.3 F 0.0
25:00 224.1 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 24:28
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:02:43
CYCLE COMPLETE

12/21/2021 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____