
Midmark M11 - v1.0.5
Total Cycles: 1954

Name: RITTER2
Sterilizer ID: V2313228
01/20/2023 11:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.8 F 0.0
2:00 77.4 F 0.0
4:00 88.4 F 0.0
6:00 117.6 F 0.0
8:00 168.8 F 0.8
10:00 200.4 F 4.3
12:00 196.8 F 8.0
14:00 209.1 F 12.0
16:00 220.5 F 6.2
18:00 233.0 F 10.6
20:00 233.0 F 8.1
22:00 242.8 F 12.9
24:00 241.2 F 11.5
26:00 253.5 F 17.7
28:00 264.6 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.2 F 29.3
2:00 272.1 F 29.8
3:00 271.5 F 28.9
4:00 272.4 F 29.5

Min 271.3 F 28.9
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 224.3 F 0.0
10:00 217.4 F 0.0
15:00 214.0 F 0.0
20:00 209.7 F 0.0
25:00 207.2 F 0.0
30:00 206.0 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 29:57
STERILIZING: 4:00
VENTING: 2:49
DRYING: 30:00
TOTAL CYCLE: 01:07:51
CYCLE COMPLETE

01/20/2023 12:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____