
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000542
11:12 09/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:42
STERILIZING 3:49
136.0 C 228kPa 3:49
136.1 C 231kPa 4:19
136.2 C 231kPa 4:49
136.3 C 231kPa 5:19
136.2 C 231kPa 5:49
136.4 C 233kPa 6:19
136.4 C 232kPa 6:49
136.4 C 233kPa 7:19
136.4 C 233kPa 7:19
VENTING 7:19
AIR DRYING 8:12
CYCLE COMPLETE 23:12

Digital Signature #
587D87401FC104E1
