

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002104  
9:24 02/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:51  
PRESSURIZING 3:00  
STERILIZING 4:08  
136.0 C 220kPa 4:08  
136.0 C 220kPa 4:38  
136.4 C 222kPa 5:08  
136.2 C 223kPa 5:38  
136.3 C 225kPa 6:08  
136.2 C 223kPa 6:38  
136.2 C 224kPa 7:08  
136.4 C 224kPa 7:38  
136.4 C 224kPa 7:38  
VENTING 7:38  
AIR DRYING 8:27  
CYCLE COMPLETE 17:45

-----  
Digital Signature #  
3B1F6AC671A1EC9D  
-----