

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3992

Name: RITTER 1  
Sterilizer ID: V2178198  
11/07/2023 04:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.4 F 0.0  
2:00 113.7 F 1.3  
4:00 131.6 F 4.1  
6:00 160.5 F 7.8  
8:00 201.6 F 12.2  
10:00 210.8 F 5.3  
12:00 222.8 F 9.8  
14:00 232.7 F 8.7  
16:00 238.5 F 11.2  
18:00 236.1 F 9.0  
20:00 250.0 F 15.5  
22:00 261.7 F 22.1  
24:00 272.6 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.5 F 29.1  
2:00 271.3 F 28.5  
3:00 272.7 F 29.2  
4:00 271.8 F 28.6

Min 271.3 F 28.4  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.1 F 0.0  
5:00 216.5 F 0.0  
10:00 219.3 F 0.0  
15:00 215.9 F 0.0  
20:00 225.7 F 0.0  
25:00 232.4 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 24:24  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:02:54  
CYCLE COMPLETE

11/07/2023 05:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_