
Midmark M11 - v1.0.5
Total Cycles: 705

Name: RITTER2
Sterilizer ID: V2313228
01/19/2022 12:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.3 F 0.0
2:00 133.7 F 1.8
4:00 155.0 F 4.8
6:00 181.6 F 9.0
8:00 200.6 F 9.4
10:00 215.7 F 7.7
12:00 230.2 F 13.0
14:00 235.1 F 10.4
16:00 234.8 F 8.5
18:00 249.3 F 15.7
20:00 261.5 F 22.7
22:00 272.8 F 30.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.5
1:00 272.4 F 29.9
2:00 271.6 F 29.7
3:00 272.6 F 29.9
4:00 271.5 F 29.5

Min 271.4 F 29.3
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 216.9 F 0.0
10:00 220.5 F 0.0
15:00 204.2 F 0.0
20:00 220.4 F 0.0
25:00 222.2 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 22:24
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

01/19/2022 01:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____