
Midmark M11 - v1.0.5
Total Cycles: 1323

Name: RITTER2
Sterilizer ID: V2313228
07/20/2022 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.0 F 0.0
2:00 163.7 F 0.0
4:00 180.5 F 1.2
6:00 207.9 F 5.8
8:00 218.5 F 11.0
10:00 224.5 F 6.6
12:00 238.4 F 12.2
14:00 237.7 F 10.3
16:00 236.5 F 9.1
18:00 251.0 F 16.2
20:00 262.5 F 22.8
22:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.4 F 29.4
2:00 271.2 F 28.9
3:00 272.5 F 29.5
4:00 271.8 F 28.9

Min 271.2 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 228.0 F 0.0
10:00 224.7 F 0.0
15:00 216.8 F 0.0
20:00 221.1 F 0.0
25:00 217.6 F 0.0
30:00 215.9 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 22:20
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

07/20/2022 11:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____