
Midmark M11 - v1.0.5
Total Cycles: 3401

Name: RITTER2
Sterilizer ID: V2313228
02/22/2024 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 71.9 F 0.0
4:00 78.8 F 0.0
6:00 105.8 F 1.2
8:00 142.4 F 4.1
10:00 171.4 F 7.1
12:00 190.6 F 10.4
14:00 209.0 F 5.6
16:00 216.4 F 7.1
18:00 226.3 F 11.0
20:00 229.2 F 7.0
22:00 239.7 F 12.0
24:00 237.9 F 10.1
26:00 250.3 F 16.1
28:00 261.2 F 22.2
30:00 271.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 29.7
2:00 271.4 F 29.2
3:00 272.6 F 29.7
4:00 271.4 F 29.1

Min 271.3 F 29.0
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.7 F 0.0
5:00 206.9 F 0.0
10:00 199.2 F 0.0
15:00 189.6 F 0.0
20:00 180.9 F 0.0
25:00 176.8 F 0.0
30:00 175.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:42
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:08:48
CYCLE COMPLETE

02/22/2024 11:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____