
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001863
14:04 16/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:06
STERILIZING 4:05
136.0 C 220kPa 4:05
136.5 C 223kPa 4:35
136.2 C 223kPa 5:05
136.2 C 224kPa 5:35
136.2 C 224kPa 6:05
136.3 C 224kPa 6:35
136.4 C 225kPa 7:05
136.1 C 223kPa 7:35
136.1 C 223kPa 7:35
VENTING 7:35
AIR DRYING 8:32
CYCLE COMPLETE 8:36

Digital Signature #
71CF1E3EAD6DD4B1
