

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2965

Name: RITTER 1  
Sterilizer ID: V2178198  
03/02/2023 09:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 127.7 F 0.0  
2:00 132.3 F 0.0  
4:00 167.9 F 0.1  
6:00 182.5 F 1.9  
8:00 202.5 F 5.3  
10:00 216.8 F 9.7  
12:00 229.1 F 8.6  
14:00 233.3 F 9.3  
16:00 241.6 F 10.2  
18:00 242.4 F 11.9  
20:00 239.8 F 10.3  
22:00 253.2 F 16.9  
24:00 264.7 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.4 F 29.0  
2:00 271.3 F 28.1  
3:00 272.4 F 29.1  
4:00 271.9 F 28.5

Min 271.2 F 28.1  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 222.0 F 0.0  
10:00 224.2 F 0.0  
15:00 215.8 F 0.0  
20:00 232.9 F 0.0  
25:00 222.3 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 25:54  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:03:58  
CYCLE COMPLETE

03/02/2023 10:05 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_