
Midmark M11 - v1.0.5
Total Cycles: 4289

Name: RITTER2
Sterilizer ID: V2313228
01/23/2025 04:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.3 F 0.0
2:00 130.3 F 1.6
4:00 152.3 F 4.5
6:00 176.6 F 8.3
8:00 194.8 F 12.4
10:00 215.6 F 4.6
12:00 227.9 F 8.5
14:00 238.2 F 12.8
16:00 234.9 F 9.4
18:00 240.8 F 10.6
20:00 245.5 F 13.4
22:00 256.3 F 19.3
24:00 266.7 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.6 F 29.8
2:00 272.4 F 29.2
3:00 271.0 F 28.9
4:00 272.0 F 29.8

Min 271.0 F 28.8
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.4 F 0.0
5:00 201.6 F 0.0
10:00 202.4 F 0.0
15:00 204.2 F 0.0
20:00 198.7 F 0.0
25:00 193.9 F 0.0
30:00 190.9 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 25:39
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:04:40
CYCLE COMPLETE

01/23/2025 05:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____