
Midmark M11 - v1.0.0
Total Cycles: 3461

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2023 02:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.5 F 0.0
2:00 185.5 F 3.6
4:00 209.5 F 8.7
6:00 221.7 F 7.9
8:00 233.9 F 10.1
10:00 234.7 F 7.9
12:00 242.6 F 11.5
14:00 246.3 F 13.4
16:00 259.2 F 20.5
18:00 271.0 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 28.9
2:00 271.7 F 28.6
3:00 272.5 F 28.9
4:00 271.5 F 28.4

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 234.0 F 0.0
10:00 229.9 F 0.0
15:00 227.0 F 0.0
20:00 226.6 F 0.0
25:00 229.2 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 18:43
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

06/28/2023 03:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____