
Midmark M11 - v1.0.0
Total Cycles: 3120

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2023 08:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.9 F 0.0
2:00 165.0 F 0.0
4:00 166.1 F 2.2
6:00 192.9 F 6.8
8:00 211.1 F 11.3
10:00 222.7 F 6.5
12:00 236.9 F 12.1
14:00 235.6 F 9.5
16:00 236.9 F 8.9
18:00 248.2 F 14.6
20:00 260.5 F 21.3
22:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.5 F 29.3
2:00 271.7 F 28.6
3:00 271.4 F 28.8
4:00 272.6 F 29.3

Min 271.1 F 28.3
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.9 F 0.0
5:00 216.4 F 0.0
10:00 238.9 F 0.0
15:00 227.3 F 0.0
20:00 222.6 F 0.0
25:00 219.0 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 22:39
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:44
CYCLE COMPLETE

04/18/2023 09:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____