
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000857
14:48 10/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 3:50
STERILIZING 5:05
136.0 C 227kPa 5:05
136.4 C 230kPa 5:35
136.2 C 229kPa 6:05
136.2 C 231kPa 6:35
136.3 C 231kPa 7:05
136.4 C 232kPa 7:35
136.3 C 232kPa 8:05
136.4 C 232kPa 8:35
136.4 C 232kPa 8:35
VENTING 8:35
AIR DRYING 9:29
CYCLE COMPLETE 24:29

Digital Signature #
2AF9DF2743450B7A
