

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3825

Name: RITTER2  
Sterilizer ID: V2313228  
07/31/2024 01:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.3 F 0.0  
2:00 135.8 F 1.4  
4:00 149.6 F 4.2  
6:00 181.2 F 8.2  
8:00 222.5 F 12.9  
10:00 223.6 F 6.5  
12:00 234.6 F 11.0  
14:00 229.5 F 7.2  
16:00 242.8 F 12.9  
18:00 240.7 F 11.3  
20:00 253.2 F 17.5  
22:00 264.0 F 24.1  
24:00 271.1 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.4 F 29.6  
2:00 271.3 F 29.4  
3:00 272.2 F 29.4  
4:00 271.6 F 29.5

Min 271.1 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.3 F 0.0  
5:00 233.0 F 0.0  
10:00 190.5 F 0.0  
15:00 185.5 F 0.0  
20:00 182.9 F 0.0  
25:00 184.4 F 0.0  
30:00 191.0 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 24:01  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:02:33  
CYCLE COMPLETE

07/31/2024 02:46 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_