
Midmark M11 - v1.0.0
Total Cycles: 3374

Name: RITTER 1
Sterilizer ID: V2178198
06/12/2023 12:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.1 F 0.0
2:00 188.4 F 4.6
4:00 206.4 F 9.4
6:00 217.6 F 6.6
8:00 229.2 F 10.2
10:00 232.6 F 7.5
12:00 243.6 F 13.2
14:00 242.6 F 11.7
16:00 255.5 F 18.3
18:00 267.0 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.3 F 28.9
2:00 271.2 F 28.5
3:00 272.3 F 29.0
4:00 271.4 F 28.3

Min 271.1 F 28.3
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 217.9 F 0.0
10:00 232.8 F 0.0
15:00 241.9 F 0.0
20:00 235.2 F 0.0
25:00 225.5 F 0.0
30:00 216.2 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 19:26
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:25
CYCLE COMPLETE

06/12/2023 01:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____