
Midmark M11 - v1.0.0
Total Cycles: 4242

Name: RITTER 1
Sterilizer ID: V2178198
01/29/2024 02:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.9 F 0.0
2:00 126.9 F 1.5
4:00 146.1 F 4.3
6:00 185.8 F 8.4
8:00 208.4 F 13.4
10:00 222.0 F 5.8
12:00 229.4 F 10.0
14:00 234.7 F 8.7
16:00 241.1 F 11.1
18:00 235.9 F 8.7
20:00 249.9 F 15.1
22:00 261.4 F 21.5
24:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 28.9
2:00 271.7 F 28.7
3:00 272.5 F 28.9
4:00 271.5 F 28.5

Min 271.4 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.7 F 0.0
5:00 217.5 F 0.0
10:00 187.5 F 0.0
15:00 182.0 F 0.0
20:00 179.0 F 0.0
25:00 229.7 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:18
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:03:45
CYCLE COMPLETE

01/29/2024 03:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____