
Midmark M11 - v1.0.0
Total Cycles: 5440

Name: RITTER 1
Sterilizer ID: V2178198
03/25/2025 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.2 F 0.0
2:00 74.5 F 0.3
4:00 104.2 F 1.7
6:00 127.5 F 4.1
8:00 155.6 F 6.8
10:00 175.1 F 9.7
12:00 192.4 F 13.3
14:00 203.6 F 3.8
16:00 215.3 F 7.2
18:00 226.9 F 10.9
20:00 229.8 F 6.8
22:00 237.3 F 10.5
24:00 237.3 F 8.2
26:00 246.3 F 13.6
28:00 257.4 F 19.6
30:00 267.5 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.9 F 29.3
2:00 271.9 F 28.5
3:00 272.8 F 29.4
4:00 272.3 F 28.8

Min 271.7 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 189.5 F 0.0
10:00 195.2 F 0.0
15:00 195.2 F 0.0
20:00 197.0 F 0.0
25:00 199.7 F 0.0
30:00 200.5 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:28
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:09:47
CYCLE COMPLETE

03/25/2025 10:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____