

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001792  
10:46 13/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:49  
PRESSURIZING 2:58  
STERILIZING 4:05  
136.4 C 222kPa 4:05  
136.1 C 223kPa 4:35  
136.4 C 224kPa 5:05  
136.2 C 223kPa 5:35  
136.4 C 225kPa 6:05  
136.4 C 224kPa 6:35  
136.4 C 223kPa 7:05  
136.3 C 223kPa 7:35  
136.3 C 223kPa 7:35  
VENTING 7:35  
AIR DRYING 8:24  
CYCLE COMPLETE 38:24  
-----

Digital Signature #  
7715969952354DE9  
-----