

-----  
Midmark M11 - v1.0.5  
Total Cycles: 270

Name: RITTER2  
Sterilizer ID: V2313228  
08/09/2021 10:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.7 F 0.0  
2:00 125.4 F 1.1  
4:00 137.5 F 3.8  
6:00 160.5 F 7.6  
8:00 189.0 F 12.1  
10:00 209.0 F 5.0  
12:00 222.7 F 9.6  
14:00 231.2 F 9.5  
16:00 237.8 F 11.5  
18:00 237.5 F 9.9  
20:00 250.0 F 16.1  
22:00 261.0 F 22.4  
24:00 271.2 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.3  
1:00 272.3 F 30.0  
2:00 272.2 F 29.5  
3:00 271.9 F 29.8  
4:00 272.7 F 29.9

Min 271.5 F 29.2  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.7 F 0.0  
5:00 209.8 F 0.0  
10:00 213.9 F 0.0  
15:00 213.5 F 0.0  
20:00 206.6 F 0.0  
25:00 201.2 F 0.0  
30:00 195.6 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 24:35  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:02:47  
CYCLE COMPLETE

08/09/2021 11:09 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01