
Midmark M11 - v1.0.0
Total Cycles: 4123

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2023 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.9 F 0.0
2:00 152.7 F 1.7
4:00 159.6 F 4.4
6:00 184.1 F 7.9
8:00 207.5 F 12.4
10:00 221.2 F 5.2
12:00 230.5 F 9.5
14:00 237.3 F 10.8
16:00 238.6 F 10.5
18:00 236.2 F 8.2
20:00 248.7 F 14.4
22:00 259.9 F 20.5
24:00 269.8 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.2 F 28.5
2:00 272.6 F 28.9
3:00 271.6 F 28.7
4:00 272.0 F 28.5

Min 271.3 F 28.3
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.3 F 0.0
5:00 196.5 F 0.0
10:00 224.2 F 0.0
15:00 197.0 F 0.0
20:00 221.1 F 0.0
25:00 215.8 F 0.0
30:00 205.2 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 25:46
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:04:35
CYCLE COMPLETE

12/15/2023 11:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____