
Midmark M11 - v1.0.5
Total Cycles: 3041

Name: RITTER2
Sterilizer ID: V2313228
10/23/2023 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.4 F 0.0
2:00 166.0 F 2.0
4:00 176.9 F 6.2
6:00 200.7 F 11.1
8:00 216.6 F 5.3
10:00 229.1 F 10.6
12:00 231.3 F 7.4
14:00 242.3 F 12.0
16:00 244.9 F 13.2
18:00 258.0 F 20.1
20:00 269.5 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.5 F 29.6
2:00 271.8 F 28.9
3:00 271.8 F 29.5
4:00 272.4 F 29.6

Min 271.2 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 212.2 F 0.0
10:00 218.9 F 0.0
15:00 211.7 F 0.0
20:00 205.7 F 0.0
25:00 203.4 F 0.0
30:00 204.3 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 21:01
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:45
CYCLE COMPLETE

10/23/2023 10:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____