

-----  
Midmark M11 - v1.0.5  
Total Cycles: 866

Name: RITTER2  
Sterilizer ID: V2313228  
03/09/2022 12:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.5 F 0.0  
2:00 193.8 F 4.3  
4:00 204.7 F 10.0  
6:00 221.4 F 6.3  
8:00 233.5 F 11.8  
10:00 236.1 F 10.0  
12:00 236.5 F 8.8  
14:00 250.3 F 16.1  
16:00 262.5 F 23.1  
18:00 272.7 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.2  
1:00 271.9 F 29.8  
2:00 272.4 F 29.5  
3:00 271.3 F 29.3  
4:00 272.4 F 30.0  
  
Min 271.2 F 29.0  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.6 F 0.0  
5:00 229.0 F 0.0  
10:00 220.5 F 0.0  
15:00 209.1 F 0.0  
20:00 215.7 F 0.0  
25:00 219.1 F 0.0  
30:00 219.5 F 0.0  
Drying Complete

FILLING: 2:23  
HEATING: 18:08  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:41  
CYCLE COMPLETE

03/09/2022 01:22 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_