
Midmark M11 - v1.0.0
Total Cycles: 2484

Name: RITTER 1
Sterilizer ID: V2178198
10/25/2022 10:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.3 F 0.0
2:00 178.7 F 3.3
4:00 193.6 F 8.2
6:00 217.5 F 7.6
8:00 229.7 F 9.6
10:00 235.9 F 8.1
12:00 243.0 F 13.3
14:00 242.7 F 12.1
16:00 256.5 F 19.3
18:00 268.5 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.7 F 29.3
2:00 271.6 F 28.7
3:00 272.7 F 29.4
4:00 272.0 F 28.7

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 210.4 F 0.0
10:00 209.5 F 0.0
15:00 210.6 F 0.0
20:00 218.0 F 0.0
25:00 226.2 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:30
CYCLE COMPLETE

10/25/2022 10:58 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____