

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2018

Name: RITTER 1  
Sterilizer ID: V2178198  
06/09/2022 01:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 159.4 F 0.0  
2:00 160.1 F 3.6  
4:00 181.3 F 8.0  
6:00 202.4 F 13.1  
8:00 220.4 F 6.9  
10:00 233.6 F 11.9  
12:00 233.8 F 8.4  
14:00 240.6 F 10.9  
16:00 244.7 F 12.7  
18:00 257.0 F 19.2  
20:00 268.0 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.4  
1:00 272.4 F 29.1  
2:00 271.5 F 28.4  
3:00 272.0 F 29.1  
4:00 271.9 F 28.8  
  
Min 271.1 F 28.3  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.9 F 0.0  
5:00 218.9 F 0.0  
10:00 208.1 F 0.0  
15:00 220.0 F 0.0  
20:00 230.6 F 0.0  
25:00 221.9 F 0.0  
30:00 212.6 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 21:17  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:54  
CYCLE COMPLETE

06/09/2022 02:32 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02