
Midmark M11 - v1.0.5
Total Cycles: 4133

Name: RITTER2
Sterilizer ID: V2313228
11/18/2024 07:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.3 F 0.0
2:00 149.0 F 0.2
4:00 175.2 F 4.1
6:00 202.7 F 9.6
8:00 218.1 F 5.6
10:00 235.3 F 10.4
12:00 232.3 F 7.5
14:00 242.0 F 11.2
16:00 246.4 F 13.7
18:00 259.5 F 20.8
20:00 270.9 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 272.7 F 29.2
2:00 272.0 F 29.1
3:00 272.5 F 29.1
4:00 272.0 F 29.2

Min 271.7 F 28.6
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.9 F 0.0
5:00 222.1 F 0.0
10:00 201.3 F 0.0
15:00 205.0 F 0.0
20:00 202.8 F 0.0
25:00 202.5 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 20:43
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

11/18/2024 07:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____