
Midmark M11 - v1.0.5
Total Cycles: 2006

Name: RITTER2
Sterilizer ID: V2313228
02/03/2023 10:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.3 F 0.0
2:00 156.0 F 0.4
4:00 165.8 F 4.1
6:00 193.8 F 9.7
8:00 216.9 F 6.2
10:00 235.4 F 12.4
12:00 239.2 F 11.0
14:00 240.0 F 10.5
16:00 255.7 F 18.5
18:00 269.0 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.5 F 28.9
2:00 272.8 F 29.5
3:00 271.7 F 28.4
4:00 273.0 F 29.3

Min 271.6 F 28.4
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 222.1 F 0.0
5:00 232.5 F 0.0
Dry Cycle Aborted

FILLING: 1:08
HEATING: 18:59
STERILIZING: 4:00
VENTING: 2:45
DRYING: 8:24
TOTAL CYCLE: 00:35:16

02/03/2023 11:18 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____