

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2184

Name: RITTER 1  
Sterilizer ID: V2178198  
07/27/2022 09:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.0 F 0.0  
2:00 162.0 F 0.7  
4:00 173.1 F 4.4  
6:00 193.8 F 8.6  
8:00 209.1 F 13.1  
10:00 220.5 F 7.1  
12:00 232.9 F 11.9  
14:00 233.3 F 8.6  
16:00 241.5 F 10.8  
18:00 244.2 F 12.6  
20:00 256.5 F 19.0  
22:00 267.3 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.5 F 29.3  
2:00 272.2 F 28.8  
3:00 271.4 F 28.4  
4:00 272.1 F 29.1

Min 271.3 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.3 F 0.0  
5:00 236.1 F 0.0  
10:00 231.0 F 0.0  
15:00 229.3 F 0.0  
20:00 230.5 F 0.0  
25:00 229.0 F 0.0  
30:00 230.5 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 23:30  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:01:30  
CYCLE COMPLETE

07/27/2022 10:19 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_