
Midmark M11 - v1.0.0
Total Cycles: 947

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2021 11:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.5 F 0.0
2:00 188.1 F 2.9
4:00 197.4 F 6.7
6:00 226.3 F 11.9
8:00 227.4 F 6.6
10:00 238.1 F 11.7
12:00 235.6 F 9.0
14:00 241.1 F 10.1
16:00 247.4 F 13.8
18:00 259.2 F 20.3
20:00 269.9 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.2
1:00 272.6 F 29.0
2:00 271.9 F 28.3
3:00 272.6 F 29.0
4:00 272.3 F 28.5

Min 271.6 F 28.1
Max 273.0 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 231.6 F 0.0
10:00 229.4 F 0.0
15:00 226.3 F 0.0
20:00 224.9 F 0.0
25:00 224.8 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 20:49
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:29
CYCLE COMPLETE

06/28/2021 12:15 PM
ACCEPTED: X
REJECTED:
Operator: U01