

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1607

Name: RITTER 1  
Sterilizer ID: V2178198  
02/08/2022 11:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.3 F 0.0  
2:00 178.3 F 3.9  
4:00 185.7 F 8.6  
6:00 208.7 F 10.8  
8:00 223.4 F 8.0  
10:00 238.8 F 13.3  
12:00 237.6 F 10.5  
14:00 236.7 F 8.5  
16:00 250.2 F 15.4  
18:00 262.1 F 22.2  
20:00 272.8 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.4 F 29.1  
2:00 271.5 F 28.3  
3:00 272.5 F 29.2  
4:00 272.1 F 28.7  
  
Min 271.3 F 28.3  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.7 F 0.0  
5:00 224.2 F 0.0  
10:00 233.0 F 0.0  
15:00 229.7 F 0.0  
20:00 227.9 F 0.0  
25:00 225.8 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 20:22  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:60  
CYCLE COMPLETE

02/08/2022 12:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_