

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2880

Name: RITTER2  
Sterilizer ID: V2313228  
09/14/2023 08:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.3 F 0.0  
2:00 74.4 F 0.0  
4:00 75.8 F 0.0  
6:00 100.9 F 0.0  
8:00 154.9 F 0.0  
10:00 192.3 F 1.2  
12:00 199.0 F 4.7  
14:00 212.9 F 7.9  
16:00 232.0 F 11.5  
18:00 227.2 F 5.8  
20:00 235.3 F 10.4  
22:00 235.2 F 8.8  
24:00 243.5 F 12.5  
26:00 240.9 F 11.0  
28:00 253.0 F 17.0  
30:00 263.5 F 23.2  
32:00 272.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.6  
1:00 271.6 F 29.1  
2:00 272.6 F 29.3  
3:00 272.3 F 29.6  
4:00 272.2 F 29.1  
  
Min 271.0 F 28.4  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.2 F 0.0  
5:00 221.8 F 0.0  
10:00 197.5 F 0.0  
15:00 177.8 F 0.0  
20:00 175.9 F 0.0  
25:00 173.9 F 0.0  
30:00 177.3 F 0.0  
  
Drying Complete

FILLING: 0:49  
HEATING: 32:08  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:10:22  
CYCLE COMPLETE

09/14/2023 09:50 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_